

AN ABSTRACT OF THE THESIS OF

Quyen L. Jones for the degree of Master of Science in Radiation Health Physics
presented on June 11, 2004.

Title: Implementation of Radiation Film Dosimetry System To Be Used For The
Verification Of a 3-D Electron Pencil-beam Algorithm On a Radiation
Treatment Planning System*

Redacted for Privacy

Abstract approved: _____

Kathryn A. Higley

Radiation film dosimetry process using the RIT 113 v.4 dosimetry software and the film digitizer VXR-12plus was used to evaluate the accuracy of electron dose calculations of the RAHD radiation therapy treatment planning system at Samaritan Regional Cancer Center. Kodak Ready-Pack EDR-2 film is recommended for dose distribution analysis in clinically practical dose ranges. The pencil-beam algorithm has a limitation for calculating dose in the penumbra region and in the tail region where the dose falls off.

* The research project was performed under the supervision of the Medical Physicist, Elizabeth A.R. Shiner, at the Samaritan Regional Cancer Center in Corvallis, Oregon.

Implementation Of Radiation Film Dosimetry System To Be Used For The
Verification Of A 3-D Electron Pencil-Beam Algorithm On A Radiation Treatment
Planning System

by
Quyen L. Jones

A THESIS

submitted to

Oregon State University

in partial fulfillment of
the requirements for the
degree of

Master of Science

Presented June 11, 2004
Commencement June 2005

Master of Science thesis of Quyen L. Jones
presented on June 11, 2004.

APPROVED:

Redacted for Privacy

Major Professor, representing Radiation Health Physics

Redacted for Privacy

Head of the Department of Nuclear Engineering and Radiation Health Physics

Redacted for Privacy

Dean of the Graduate School

I understand that my thesis will become part of the permanent collection of Oregon State University libraries. My signature below authorizes release of my thesis to any reader upon request.

Redacted for Privacy

Quyen L. Jones, Author

TABLE OF CONTENTS

	<u>Page</u>
1. Introduction.....	1
2. Background.....	3
3. Methods and Materials.....	5
3.1. Film Dosimetry.....	5
3.2. RIT 113 v.4 Film Dosimetry System Set Up.....	9
3.3. Radiation Treatment Planning System RAHD.....	10
3.4. Solid water phantoms used for the pencil-beam algorithm verification....	16
4. Results and Discussion.....	19
4.1. OD Calibration.....	19
4.2. Dose calibration.....	21
4.3. Relative depth dose analysis in film dosimetry.....	25
4.4. Comparing iso-dose distribution calculated from RAHD and doses measured using the RIT 113 system.....	32
5. Conclusion.....	37
6. Bibliography.....	39
7. Appendices.....	41

LIST OF FIGURES

<u>Figure</u>	<u>Page</u>
1. Figure 3.1 Film digitizer Vidar-VXR model.....	6
2. Figure 3.2 Linear accelerator Varian 2100C used in the experiment and film positioned inside a solid phantom in which the film-plane is parallel to radiation beam.....	6
3. Figure 3.3 Schematic view of set-up for a 2-D pencil beam algorithm for calculation of arc electron dose distributions modeled by Hogstrom et al.....	14
4. Figure 3.4 Schematic of irradiation geometry of HDPE phantoms.....	17
5. Figure 4.1 Optical density corresponding to scanner signal values (A/D) from Vidar VXR-12P Film Digitizer and the RIT 113 V4 Film Dosimetry Software.....	19
6. Figure 4.2 Relationship between scanner's A/D signal values to dose.....	22
7. Figure 4.3 Plot of $\ln[\log(S_0/S)]$ versus $\ln(D)$ yield a straight line with slope α and intercepted at $\ln(a)$	23
8. Figure 4.4 Optical density versus relative dose on XV-2 and EDR-2 films.....	24
9. Figure 4.5a-e Optical density versus relative percent depth dose curves on EDR-2 films exposed with 150MU of electron radiation at various energy levels using a typical cone size.....	27
10. Figure 4.6a-e Relative percent depth dose versus scanner signal curves on EDR-2 films exposed with 150MU electron radiation using various cones with different sizes.....	31
11. Figure 4.7a-e Depth iso-dose distribution in polystyrene phantom from linear accelerator Varian 2100C of five electron energy levels at an SSD of 100cm using 15cm x 15cm field cone.....	34

LIST OF APPENDICES

<u>Appendix</u>	<u>Page</u>
A. Optical Density Calibration File with the associated OD versus Scanner A/D Values plot and the Stouffer Step Wedge Image.....	42
B. Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 150MU_ 4cm x 4cm cone _ 6MeV, 9MeV, 12MeV, 16MeV, and 20MeV.....	46
C. Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 150MU_ 6cm x 6cm cone _ 6MeV, 9MeV, 12MeV, 16MeV, and 20MeV.....	57
D. Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 150MU_ 10cm x 10cm cone _ 6MeV, 9MeV, 12MeV, 16MeV, and 20MeV.....	68
E. Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 150MU_ 15cm x 15cm cone _ 6MeV, 9MeV, 12MeV, 16MeV, and 20MeV.....	79
F. Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 150MU_ 20cm x 20cm cone _ 6MeV, 9MeV, 12MeV, 16MeV, and 20MeV.....	90
G. Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 100MU_ 15cm x 15cm cone _ 12MeV.....	101
H. Parallel dose calibration file using Kodak Ready-Pack XV-2 film _ 100MU_ 15cm x 15cm cone _ 12MeV.....	104
I. Parallel dose calibration file using Kodak Ready-Pack XV-2 film _ 50MU_ 10cm x 10cm cone _ 12MeV.....	107
J. Film image and iso-dose contours on Kodak Ready-Pack EDR-2 film using 150MU_ 15cm x 15cm cone _ 12MeV.....	110

LIST OF APPENDIX FIGURES

<u>Figure</u>	<u>Page</u>
1. Figure A.1 Plot generated from the OD calibration file by RIT 113.v4 software.....	45
2. Figure A.2 Stouffer Step Wedge Film Image has 32 steps (32 shades of gray).....	45
3. Figure J.1 6MeV 15x15 cone 150MU EDR film	111
4. Figure J.2 9MeV 15x15 cone 150MU EDR film	112
5. Figure J.3 12MeV 15x15 cone 150MU EDR film	113
6. Figure J.4 16MeV 15x15 cone 150MU EDR film	114
7. Figure J.5 20MeV 15x15 cone 150MU EDR film	115

1. INTRODUCTION

The purpose of this thesis is to evaluate the accuracy of the electron dose calculations of the RAHD radiation therapy treatment planning system at the Samaritan Regional Cancer Center. The central core of this thesis is to initiate and implement a film dosimetry process which is used for the verification of the electron beam dose distributions as calculated by the RAHD's electron pencil beam algorithm and modify the algorithm parameters as necessary to match the linear accelerator output. A set of phantoms was also developed for the study of the accuracy of the RAHD system in complex treatments that utilize a combination of electron energies with irregular field shapes in inhomogeneous medium.

RIT 113 (version 4) softwareⁱ, in conjunction with a film digitizer Vidar VXR-12 *plus*, was calibrated and used to analyze films that were exposed to different electron beams with different field (cone) sizes. The RIT software created isodose distributions from the films that were then compared to the isodoses from the RAHD treatment planning system. Information on the linear accelerator's beam characteristics which was used as input for RIT 113 film dosimetry and RAHD treatment planning system are provided by the blue water phantom system in association with the OmniPro-Accept software platformⁱⁱ. A set of humanoid phantoms was developed to assess the accuracy of RAHD's dose calculation in

ⁱ Radiological Imaging Technology, Inc. , Colorado Springs, CO Tel. 719-590-1077

ⁱⁱ Scanditronix Wellhofer North America, Bartlett, TN Tel. 901-386-2242

complex field shapes and surface irregularities. Once the algorithm parameters were successfully modeled within the RAHD, phantoms were scanned with the CT scanner and networked into the RAHD system. From this, more complex electron dose distributions measured by the film dosimetry process and calculated by the treatment planning system were compared. Kodak Ready-Pack XV-2 films and EDR-2 filmsⁱⁱⁱ were used in the film dosimetry process; their features in responding to electron irradiations were also investigated.

ⁱⁱⁱ Eastman Kodak Company – Health Imaging Products, Tel. 800-328-2910

2. BACKGROUND

At the Samaritan Regional Cancer Center in Corvallis, Oregon, external beam radiation therapy is used as a means of treating cancer patients. It is currently the most common means of external radiotherapy and involves the application of an electron beam(s) and/or a high energy X-ray(s) produced by a linear accelerator. At the center, a Varian-2100C linear accelerator is used. The Varian-2100C generates either X-rays of 6MV and 18MV or electron beam radiations of 6, 9, 12, 16, and 20 MeV which are precisely directed toward and into a target volume treatment.

RAHD is used at the cancer center for setting up a treatment plan according to a prescription from a radiation oncologist. Because dosimetrists and radiation oncologists rely on RAHD for precise treatment plans designed to deliver high doses to target tumor volumes while sparing the surrounding normal tissue as much as possible, verification of the treatment planning system is vital for ensuring that it accurately models the accelerator output. The radiation treatment planning system must be both accurate and precise in terms of the dosage used and the spatial geometry of the treatment delivery, to within $\pm 2-3\%$ of that required [17, 18]. This requirement of accuracy and precision makes quality assurance (QA) a major consideration of any treatment planning system. Small deviations from the actual dosage and spatial geometry of the accelerator can result in harm to healthy tissue, or, a non-effective dose to the tumor. Improving the accuracy and precision of the

method of dose calculations therefore not only helps to prevent a detrimental effect to the surrounding healthy tissue, but to improve cancer cure rates. The reported dose must be consistent with existing protocols, i.e., methods of dose specification [17, 18]. In other words, there must be assurance that the calculated dose from the treatment planning system is matched with the measured dose with a deviation within the tolerance.

RAHD calculates plan doses in both 2-D and 3-D utilizing a pencil-beam algorithm. The parameters for the pencil beam algorithm effectively influence the accuracy of the calculated doses output [3]. A film dosimetry process was initiated and implemented to evaluate the accuracy of the algorithm used in the treatment planning system. RIT 113 (version 4) software was used in the film dosimetry process. It is the most advanced film dosimetry available and is widely used in clinical settings. Materials, procedures, and possible preferences in the film dosimetry process were explored during the initiation of the project.

3. METHODS AND MATERIALS

3.1 *Film dosimetry*

Film dosimetry can be used to assist in the verification and modification of the radiation treatment planning system. The iso-dose distributions calculated by the radiation treatment planning system can be compared to actual iso-dose distributions output from the linear accelerator as viewed on an exposed film. In essence, film dosimetry QA can be used to modify the treatment planning system via a “feedback” method by which the parameters of the pencil-beam algorithm (of the RAHD) can be adjusted until the calculated and actual dose distributions match to within the required specifications. Film dosimetry is well established and widely used for determining relative electron beam dose distributions. Film dosimetry is favored for its high spatial resolution and high density/dose resolution that gives the medical physicists the ability to perform fast and accurate dose analysis in numerous clinical QA situations such as multi-leaf collimator test, IMRT, dynamic wedge, and stereotactic radiosurgery. Depending on the technology employed in the process, the range of the integrated spatial dose distribution can be from just a few cGy to several hundreds cGy. In addition, radiographic films allow repetitive readouts and provide a permanent record of the dosimetric measurement. Other advantages include the flexibility to place film in a humanoid phantom, short measurement time, and intrinsically two dimensional measurements [2, 15, 17].

The most common technique in film dosimetry is to sandwich a film within a plastic solid water phantom and expose the film to a radiation beam with the film-plane either parallel or perpendicular to the central axis of the radiation field [2, 8, 15]. Figure 3.1 shows the Vidar-VXR model film digitizer and the Varian 2100C linear accelerator used on the experiment is shown in Figure 3.2. The gantry and table are set at 90° from home position. Film is sandwiched between high density polyethylene phantoms and irradiated with an electron beam parallel to the film-plan.




Figure 3.1 Film digitizer Vidar-VXR model




Figure 3.2 Linear accelerator Varian 2100C used in the experiment and film positioned inside a solid phantom in which the film-plane is parallel to radiation beam

Photographic film consists of a transparent polyester base and is coated with a radiation-sensitive emulsion which is covered with a protective layer of gelatin to keep it safe from mechanical damage. The emulsion is made of silver halide crystals; typically 95% of silver bromide and 5% of silver iodide, embedded in gelatin. The exact composition of the emulsion varies by manufacturer and types of film. When exposed to a radiation field, the silver halide forms latent images (darkened regions) in the film due to ionization and excitation. Film development is a chemical process that amplifies the latent images by a factor of millions. The dose associated with the quantity of the radiation can be related to the darkening of the film and thus depends on the film emulsion. Some emulsions are fast and require a small amount of radiation to darken the film. Other emulsions are slow and required higher doses to produce the same darkness. Thus, the types of film are chosen depending on application. These darkened regions can be related to the intensity of the incident radiation. The light and dark regions on an exposed film can be quantified by measuring the light transmitted through the film, characterized by an optical density unit at the point of interest on the film. The optical density is measured using a densitometer which consists of a light source on one side of the film and a receptor on the other side to measure the light transmitted. After going through chemical processing, even an unexposed film will attenuate some incident light and is regarded as the background base on the film dosimetry process. This is referred to as fog and produces the film scanner's signal

called “dark current”. The net optical density, OD, of a point on the film is defined as:

$$OD = a \log_{10} \left(\frac{S_o}{S} \right) \quad (3.1)$$

where S_o is the scanner signal for the “dark current” and S is the scanner signal for the transmitted light on the film at the point of interest. The optical density depends on the dose by the following equation:

$$\log_{10} \left(\frac{S_o}{S} \right) = a D^b \quad (3.2)$$

where a and b are constants. Taking the nature logarithm both sides of equation (3.2) results in a linear relationship:

$$\ln \left[\log_{10} \left(\frac{S_o}{S} \right) \right] = \ln(a) + b \ln(D) \quad (3.3)$$

A plot of $\ln \left[\log_{10} \left(\frac{S_o}{S} \right) \right]$ against $\ln(D)$ generates a straight line with slope b and intercept at $\ln(a)$ [15, 17].

To use a film for dosimetry, it is necessary to know the relationship between dose, optical density, and scanner signal. Some emulsions produce a linear relationship between dose and OD, which is the most favorable case, some are relatively linear over a limited range, and others are non-linear. To start with, a set of OD calibration files is created to correlate the relationship between the scanner signals and an optical density. A film digitizer (scanner) with higher resolution can

distinguish a higher range of optical density. A step wedge film composed of bands with graduated shades of gray (up to 32 steps on the Stouffer Step Wedge film) in which the average OD for each shade known is used to establish the OD calibration file. The film digitizer scans the step wedge film and generates a single scanner signal for each density and the RIT software then creates the OD calibration file. Another set of dose calibration files is required to convert the scanner signal of a corresponding OD to a dose at any point of interest. This is done by exposing several films to known radiation doses at specified locations on the films. A plot of net OD against dose, called an H&D curve (after Hurter and Driffied who originated sensitometry) is generated by the software using the OD calibration file and the inputs from the above films. This curve should be independent of the dose rate, energy, and radiation field sizes [15]. The dose calibration file is then used to analyze dose distributions on films.

3.2 RIT 113 v.4 film dosimetry system set up

The film dosimetry process is complex, requiring sophisticated scanning equipment and software. The RIT113 system consists of hardware and software for film analysis. The hardware configuration includes a film digitizer, a computer system, and a printer. The software consists of an image processing platform and various application modules. The exposed films are first scanned in the film digitizer and the images are analyzed with a film analyzing platform software package, converting the scanner signal of a sequence of pixels to optical density

units which are then converted to dose. Information for the conversion is acquired from the OD calibration file and the dose calibration file.

At the Samaritan Regional Cancer Center, the software used for radiation film dosimetry is RIT113.version 4, a sophisticated software package widely used for analyzing films. The software is capable of creating depth dose profiles, cross-section analysis, and isodose contouring, as well as generating meaningful data in advanced QA routines such as star shot analysis, radiation/light field coincidence, IMRT treatments, stereotactic plan analysis, MLC QA, patient positioning, interrupted treatment analysis, and asymmetric field analysis. The film digitizer used was a Vidar VXR-12plus. The digitizer operates with 12 bit mode of density resolution, has maximum spatial resolution of 85 μ m, and is capable of distinguishing 4,096 shades of gray. Other available spatial resolutions are 85, 89, 169, 178, 339, 356, and 423 μ m.

The spatial resolution used in this experiment was 169 μ m which is the digitizer's optimum operating resolution. The scanner's spatial resolution was calibrated by adjusting the scanner pixel size to match a measured and scanned value using the *scanner spatial calibration* routine from RIT113 v4.

3.3 Radiation Treatment Planning System RAHD

At Samaritan Regional Cancer Center, the RAHD^(TM) treatment planning system is used to create a radiation treatment plan based on a prescription from the radiation oncologist. It is a three dimensional, computerized, radiation treatment

planning system with sophisticated dose calculation algorithms allowing accurate treatment plans that target cancerous tumors while sparing the surrounding healthy tissue. The RAHD radiation treatment planning (RTP) system uses a CAT scan (computerized axial tomography) to create a three dimensional model of the patient's critical structures and the areas of interest are outlined with the contours in CT images. RAHD generates the treatment volumes with the precise locations of the area to be treated in three dimensions. The images are displayed simultaneously in the same window with all the necessary information for beam placement, including 3D surface rendering and visualization with sagittal, coronal, transverse projections, or Beam's Eye Views (BEV) with Multi-Leaf Collimator display, transverse as well as reformatted image options, and Digitally Reconstructed Radiographs (DRR) of the BEV. Because the tumor's volume and location are mapped in three dimensions, the path of the radiation beam can be better targeted toward the tumor, which is especially important when cancerous organs are located in sensitive or complex body areas, such as the prostate, brain, breast, or esophagus.

The RAHD system employs a three-dimensional pencil-beam algorithm developed through the work of Dr. Dennis Leavitt, PhD. at the University of Utah, for calculating electron beam dose distributions. The pencil-beam dose calculation algorithm allows the accurate calculation of heterogeneity effects resulting from irregular external contours or smaller internal heterogeneities. In addition to the

beam characteristic profile data, which is measured in water phantom with a NIST calibrated ion chamber system, RAHD requires certain electron beam parameters in order to accurately model the electron beams and calculate the resulted dose distributions (according to the RAHD user's operating menu book). The following parameters are energy dependent and can be modified using a built-in *utility* function within the RAHD-RTP system:

- Energy – mean energy of electrons incident upon the phantom, determined by the formula: $E = 1.919(R_p) + 0.722$;
- Practical Range R_p – determined by extrapolating the depth dose curve to the intersection with the Bremsstrahlung contribution which should be field size independent;
- Drift space – distance from the downstream edge of the beam defining collimator and the standard treatment surface (for a Varian 2100C linear accelerator the drift space is 5cm);
- Sigma-Theta-X (σ_x) – sigma of the Gaussian distribution of the electron angles (projected onto a planar surface parallel to the beam) for those electrons comprising a pencil beam defined at the plane of the beam defining collimator;
- Scatter Correction Factor – the multiplication factor applied to the linear scattering power of the reference medium (water) in the calculation, typical

adjustment range of 1.0 – 1.4, giving an increase in the penumbra of about 15%;

- Added Margin Width – accounts for the scattering of electrons well beyond the regions immediately accounted for by the pencil-beam algorithm.

Besides the above parameters which are accounted for the electron beam's characteristics, the pencil-beam algorithm integrated in the RAHD-RTP system also requires other energy-dependent correction factors to ensure the calculated depth dose closely matches the measured depth dose in water [3, 4]. The correction factors were determined by the fine-tune method in which the calculated depth doses were forced to fit the measured depth dose in the penumbra region by adjusting the correction factors optimally. These energy-dependent correction factors are:

- SIGC – Sigma for Edge Correction ;
- AZERO – Magnitude of Edge Correction;
- ALAMBDA – Exponential Depth Dependence for Profile;
- FMCS @ R_p (FMCSR_P); and
- PKX0 – Field Size at which the conical x-ray background profile goes to zero.

Pencil-beam algorithm calculation models are derived from Fermi-Eyges theory of multiple-Coulomb-scattering [1]. Figure 3.3 shows the schematic view of the set-

up for a 2-D pencil-beam algorithm for calculation of the arc electron dose distributions modeled by Hogstrom et al [1, 11]. This is based on the pencil-beam




Figure 3.3 Schematic view of set-up for a 2-D pencil beam algorithm for calculation of arc electron dose distributions modeled by Hogstrom et al. The secondary collimator can be irregular shape. The irradiated area defined by patient collimation is modeled as a single broad beam comprised of a set of pencil beams, which are summed to form strip beams for the dose calculation. This 2-D pencil beam model based on the Fermi-Eyges small angle scattering theory assumed the variations in tissue density outside the plane of calculation to be the same as in the plane of calculation.

approach in which the arc beam is designated as a single broad beam defined by the irradiated patient surface within the skin collimation; the broad beam is constructed from many pencil beams which are summed in the Y-direction to form strip beams as illustrated in Figure 3.3. The electron planar fluence passing through the secondary collimator is independent of the X-coordinate and varies slowly in the Y-direction. Each strip beam is described by an electron planar fluence and an angular distribution. The angular distribution is modeled as Gaussian, with the same mean projected angle and root-mean-square (RMS) spread about the mean projected angle as the actual distribution. By knowing the dose distribution from each strip beam in the plane of calculation of the patient, the dose at any point can be determined by performing a sum of the dose contributions from each strip beam. The dose calculations are performed in planes perpendicular to the isocenter axis. The calculation is repeated in multiple planes to form a three-dimensional dose distribution. Variations over the incident beam, changes due to surface shape and irregular fields, and effects due to inhomogeneities can be taken into account for each pencil beam individually in order to calculate dose distributions. One limitation of the pencil-beam algorithm approach is the inability to predict the off-axis ratios accurately. This is because Fermi-Eyges theory is a small-angle scattering theory and is dealt with along the central ray of each pencil beam only, while neglecting large-angle scattering arising from the electron-electron interactions. Therefore, effects due to narrow inhomogeneities are not predicted accurately that discrepancies of up to about ten percent were reported

predominantly in the penumbra region where the dose spreads significantly at the tails of off-axis dose profile [1, 3, 4, 7, 10, 11, 17]. For that reason, the user must understand the details and limitations of the pencil beam approach on the dose calculated from a radiation treatment planning system.

3.4 Solid water phantoms used for the pencil-beam algorithm verification

A set of phantoms was developed for the study of the accuracy of the RAHD-RTP system in complex treatments that utilize a combination of electron energies with irregular field shapes in inhomogeneous medium. In order to verify the performance of the treatment planning system in calculating doses of many clinical treatment situations the phantoms were constructed in different regular and irregular shapes. Dose calculations in complex inhomogeneities were also considered in assembling the phantoms [1, 6, 12, 13, 16]. Because polystyrene phantoms are used for the routine linear accelerator QA at the facility, the phantoms were mainly used in the experiment. The polystyrene phantoms have surface areas as 25cm x 25cm and various thicknesses with the thickest width was 2.8cm.

High-density polyethylene phantoms with specific density approximately 0.95g/cc were used to evaluate the algorithm in calculating dose distributions in treatment situation that model clinical situations [12]. Figure 3.4 illustrates the dimensions of the phantoms. The phantoms are black to prevent light leakage from causing film fog which would interfere with the accuracy of the electron

radiation doses. The surfaces which interface with the inserted films are smooth and polish to avoid mechanical artifact caused by roughness. The phantoms were designed to be convertible between regular shapes to various irregular shapes by switching the beam interacting-surfaces to a desired surface. In addition, the phantoms offer the options of treated medium in homogenous or in diverse inhomogenous by filling the gap with different materials such as paraffin, cork, Styrofoam, or air.




Figure 3.4 Schematic of irradiation geometry of HDPE phantoms.

Although high density polyethylene (HDPE) is not ideal water equivalent, but the density and mass-stopping power corrections to water can be acquired by the use of Computed Tomography. Computed Tomography can differentiate between tissues of similar density (measured in Hounsfield units) and provide detailed density

information which can be used for correcting dose calculation in water and other medium [6].

Regrettably, the set of HDPE phantoms were not used in the experiment to evaluate the performance of the RAHD system in calculating electron dose distributions in complex treatments. Further work is needed to accomplish the verification of the algorithm in three-dimensional dose calculations utilizing these phantoms.

4. RESULTS AND DISCUSSION

4.1 OD calibration

To determine the OD calibration curve, a Stouffer Step Wedge film, provided by Radiological Imaging Technology, is used in the *step wedge calibration* process. There are 32 steps (32 shades of gray) in this step wedge film with a single OD value for each of the steps. These OD measurements were calibrated to a NIST standard. Figure 4.1 shows the curve of OD versus A/D scanner (scanner signal value) generated by Excel based on the data from the OD calibration file. Note that RIT113 generated the same curve in BITM format. The OD calibration file with data in word format can be found in Appendix A. This OD




Figure 4.1 Optical density corresponding to scanner signal values (A/D) from Vidar VXR-12P Film Digitizer and the RIT113V4 Film Dosimetry Software

calibration file was used to develop all of the dose calibration files required for the project film analysis. The scanner signal value (S) of a pixel is related to its OD by the following equation:

$$OD = -0.3793\ln(S) + 3.206 \quad (3.4)$$

with a logarithm regression coefficient $R^2 = 0.9998$.

Equation 3.4 can be rewritten as:

$$OD = -0.3793\left(\frac{\log(S)}{\log(e)}\right) + 3.206 \quad (3.5)$$

$$= \frac{-0.3793}{0.4343} \log(S) + 3.206$$

$$= -0.8734\log(S) + 3.206$$

$$= 0.8734[-\log(S) + 3.6708]$$

$$= 0.8734[-\log(S) + \log(4686.364)]$$

or

$$OD = 0.8734\log\left(\frac{4686.364}{S}\right) \quad (3.6)$$

which is in the form of the optical absorption law.

The above plot indicates that this film scanner does not have the resolution capability to distinguish doses of corresponding optical densities above 2.5; which is the maximum optical density the scanner is able to make a distinction in a range of gray shades. The optical density versus scanner signal's curve can be divided into three regions:

- $OD < 0.85$: the scanner signals change dramatically with a small change in optical density and the scanner signal is approximately linear with optical density;
- $0.84 < OD < 1.44$: non-linear region; and
- $OD > 1.44$: the scanner signal is not sensitive with even dramatic changes in optical density and the scanner signal is approximately linear with optical density.

Therefore, for the VXR-12plus, dose analysis will be most accurate when using film with an exposure dose that produces gray shades with optical densities less than 0.85.

4.2 Dose calibration

Two types of Kodak Ready-Pack films were used in the experiment: XV-2 and EDR-2 films. Because the composition of the emulsions is different between these two types of film, their responses to the same quantity of an electron radiation are different. In other words, with the same irradiated electron dose, the resulting optical density corresponding to that dose is different for the two types of film and thus, the scanner produces different signals for the XV-2 and EDR-2 film.

A water phantom system (Wellhoffer Blue phantom tank) was utilized to collect data representing electron beam's characteristics for each electron energy and cone size (field aperture). The data provides information on the distribution of the deposited electron doses in water including depth dose profile, iso-dose

contours, cross-plane dose profile, mean energy, and practical range which were used to determine the electron parameters for the pencil-beam algorithm for each electron energy. The information on depth dose profile for each energy level with every cone size was used in establishing dose calibration files for the film dosimetry process. An exposed film can be analyzed by using the associated dose calibration file with identical electron energy and cone size. Figure 4.2 shows the relationship between the scanner's signal A/D value and dose from EDR-2 film for a 12MeV electron beam using 150MU and the 10cm x 10cm cone.




Figure 4.2 Relationship between scanner's A/D signal values to dose. The data is from EDR-2 film exposed with 150MU using 12MeV electron beam with 10cm x 10cm cone.

The correlation equation for the dose (D) versus scanner signal value (S) curve is:

$$D = -55.684 \cdot \ln(S) + 453.64; \quad R^2 = 0.9913 \quad (4.1)$$

$$\begin{aligned}
 \text{or } D &= -55.684 \left[\frac{\log(S)}{\log(e)} \right] + 453.64 \\
 &= \frac{55.684}{0.4343} \left[-\log(S) + \frac{453.64}{0.4343} \right] \\
 &= 128.217 * \log\left(\frac{3451.917}{S}\right)
 \end{aligned}$$

$$\text{Thus, } D = 128.217 * \log\left(\frac{3451.917}{S}\right); \quad R^2 = 0.9913 \quad (4.2)$$




Figure 4.3 Plot of $\ln[\log(S_0/S)]$ versus $\ln(D)$ yield a straight line with slope $_$ and intercepted at $\ln(a)$. The scanner does not respond linearly with dose when the dose dropped below 5% of D_{max} ($\sim 7.5\text{cGy}$)

S_0 value of 3451.917 is given by the equation 4.2. A plot of $\ln[\log(S_0/S)]$ versus $\ln(D)$ yield a straight line (see Fig. 4.3) with the linear regression coefficient $R^2 = 0.9947$ as discussed in the equation 3.3. The slope of the line is 1.1139 and the intercept at negative 5.4463 gives a value of “a” equal 4.312E-03. Therefore, for

each signal acquired from the film digitizer, the corresponded dose in this case can be determined as:

$$\log\left(\frac{S_0}{S}\right) = 4.312E-03 * D^{1.1139}; \quad R^2 = 0.9947 \quad (4.3)$$

The abrupt curve occurs at low dose regions (below 5% D_{max} which is approximately 7.5cGy in this case) reflecting the differences of the silver halides in responding to the electron beam in the penumbra region.




Figure 4.4 Optical density versus relative dose on XV-2 and EDR-2 films. The films were exposed with 12 MeV electron radiations at various doses on 15cmx15cm field at an SSD of 100cm. Each MU (monitor unit) was calibrated to 1cGy on the linear accelerator.

Figure 4.4 presents optical densities corresponding to doses at several dose levels on the two types of films. Optical density increased slowly and approximately linear with dose in the 0-150cGy dose range in EDR-2 films. On the contrary, optical density increased rapidly with dose in XV-2 films and became

saturated around 100cGy. The reason of saturation is the film scanner VXR-12plus does not have the resolution capability to distinguish dose higher than 100cGy where the produced optical density is about 2.5 if XV-2 films were used for dose analysis. For this limitation, EDR-2 film is recommended in high electron dose analysis. EDR-2 film is also recommended for its emulsion's response to dose in such the way that the produced optical densities were nearly below the non-linear region to scanner signals while the scanner is able to differentiate small changes of optical density in a wider range of dose. Another advantage of EDR-2 films is that they are less sensitive to chemical processing than XV-2 films. The deviation of doses, shown as the degree of change on the slopes of the lines, caused by variations of processing is stronger in XV-2 films comparing to EDR-2 films as illustrated in Figure 4.4.


4.3 Relative depth dose analysis in film dosimetry

Data was collected in exposing the polystyrene phantom with film sandwiched in between using various cone sizes and various electron energies at 150MU to analyze the relative depth doses on EDR-2 films. All the films came from the same batch and were processed in the same processor with small time space (few minutes) in between exposing and processing. Figures 4.5a-e illustrates the effect of energy levels on the dose responded on the films when the same cone size was used. For small cones, sizes of 4cm x 4cm and 6cm x 6cm the variation


due to different electron energy level was greater than for larger cones. However, for the 4cm x 4cm cone the deviation was not very significant except at 6MeV.

Figures 4.6a-e illustrates the effect of cone size on the dose for each electron energy. The 9MeV and 20MeV curves overlap for the relative doses and the size of the cones does not affect the deposited dose significantly. For the 6MeV curves, the 4cm x 4cm and the 6cm x 6cm cones' curves overlap, however, the 10cm x 10cm, 15cm x 15cm, and 20cm x 20cm cones' curves also overlap but in a different region in the plot. In contrast, the deposited doses for the 12MeV and 16MeV differ across all cone sizes.


The reason for the discrepancies between the deposited doses when comparing the energy levels and cone sizes is due to the firmware parameters of the linear accelerator. The generated figures can be used in accelerator QA to monitor the performances in producing consistent electron doses in treated volumes routinely if it is desirable. It would be nice to be able to make predictions for a dose distribution based on cone size and beam energy. However, based on this data set it is difficult to predict accurately the relative dose distribution based on cone size and electron beam energy due to the fact there is so much variations contributing in this data set. For example, variation in temperature and chemical concentration during processing can influence the dose versus optical density relationship. For that reason, a set of dose calibration file at each energy level and cone size needs to be created for each dose analysis at the same energy and cone size.




(a) 4cm x 4cm
cone




(b) 6cm x 6cm
cone



(c) 10cm x 10cm
cone




(d) 15cm x 15cm
cone




(e) 20cm x 20cm cone


Figure 4.5a-e Optical density versus relative percent depth dose curves on EDR-2 films exposed with 150MU of electron radiation at various energy levels using a typical cone size. The plots demonstrate the effect of electron energies on the deposited doses.




(a) 6MeV




(b) 9MeV



(c) 12MeV



(d) 16MeV





(e) 20MeV

Figure 4.6a-e Relative percent depth dose versus scanner signal curves on EDR-2 films exposed with 150MU electron radiation using various cones with different sizes. The plots demonstrate the effect of cone size on the deposited doses.


4.4 Comparing iso-dose distribution calculated from RAHD and doses measured using the RIT113 system

Various electron energy levels were used with a 15cm x 15cm size cone to study the iso-dose distributions on a solid polystyrene phantom. Two blocks of the polystyrene phantom (25cm x 25cm x 2.8cm), with a Kodak ready-pack EDR film sandwiched in between, were scanned in the CT scanner and the generated three-dimension image was transferred into the treatment planning system for dose scenario settings and calculations. The iso-dose distributions created from the two systems are presented in one plot for comparison as shown in Figure 4.7a-e. The iso-dose distributions within the areas of 12cm x 12cm, with iso-centers lying in the central axis of the beam, are in agreement between the two systems. Beyond this region, along the edges parallel to the central axis of the beam and at extended







(b) 9MeV



(c) 12MeV



(d) 16MeV



(e) 20MeV

Figure 4.7a-e Depth iso-dose distributions in polystyrene phantom from linear accelerator Varian 2100C of five electron energy levels at an SSD of 100cm using 15cm x 15cm field cone. Kodak ready-pack EDR films were used in the experiments. The color contours are calculated depth iso-doses, generated from the treatment planning system RADH and the black contours are measured depth ios-doses on films, generated from the film dosimetry system RIT113. (a) 6MeV (b) 9MeV (c) 12MeV (d) 16MeV (e) 20MeV

depth doses of below 5%, the dose distributions became inconsistent between the two systems. Discrepancies up to about 10% were observed for the 16MeV in this region; however the difference was less extreme for the other electron energy beam. Disagreement of within 1mm was perceived for the 20MeV. Because of the inaccuracy of the pencil-beam algorithm in the penumbra region, as discussed previously, discrepancies between the calculated doses and the measured doses are expected in this area. Another factor that contributes to dose discrepancies occurs when the electrons slow down (as when moving through a dense medium due to collisions), lowering the energy of the electrons in the penumbra region. Here, the specific ionization (the number of ion pairs formed per unit distance traveled by the electrons) increases rapidly which is also strongly affected by the difference between the Z (atomic number) of the silver halide in the film and the polystyrene. Because of the higher Z number of the silver halide the dose falls off more slowly in the film than in the polystyrene. Another effect contributing to the discrepancy above the D_{max} region is termed bremsstrahlung where X-rays are emitted when the high energy electrons undergo rapid acceleration in the region of a nucleus. Bremsstrahlung is dependent on Z as given by:

$$f_e = 1 \times 10^{-3} Z E \quad (4.3)$$

where f_e is the fraction of the energy in the electron beam that is converted into X-rays, Z is the atomic number of the absorber, and E is the energy of the electrons in MeV [5]. Therefore, a higher fraction of the energy of the electron beam is

converted to X-rays in the silver halide film than in the polystyrene. It is difficult to calculate the intensity of bremsstrahlung as well as the shape of the bremsstrahlung spectrum but easy to measure.

5. CONCLUSION

Radiation film dosimetry is an effective method in verifying the accuracy of dose distribution calculations of the radiation treatment planning system. Kodak Ready-Pack EDR-2 film is recommended for producing a high optical density to dose resolution. A wide range of doses with responsive range up to 400cGy according to the Kodak manufacture which includes clinically practical dose ranges can be achieved in dose distribution analysis. Data providing information on electron beam characteristics plays a key input for generating appropriate pencil-beam parameters for the RAHD system, as well as in creating precise dose calibration files in the film dosimetry. For that reason, collecting data on the beam's characteristics which employs the Blue-Water phantom system should be performed properly in order to avoid errors in the verification of the treatment planning system. The beam data used in a treatment planning system is always collected using a calibrated ion chamber and a water phantom system. All beam calibrations are referenced to water as the "gold standard." However, this data is very limited in its application to treatment planning (i.e. standard, square, electrons are rarely used in cancer treatment). Therefore it is necessary for the treatment planning system to accurately model the electron beam mathematically for it to have legitimate clinical use, such as irregularly shaped fields and tissue inhomogeneity.

From this study, it is shown that the pencil-beam algorithm has a weakness for calculating dose in the penumbra region and in the tail region where the dose fall off. If a cone size 15cm x 15cm is used, the measured central doses within 0.5cm of the field edge are almost matched within 1mm or 2% with the calculated doses. Beyond this region and extended to where the depth dose is below 5%, discrepancies up to 10% were observed for the 16MeV electron beam. For other electron energies (6, 9, 12, and 20MeV) the discrepancies are reduced to less than 5% or within 2mm.

In order to achieve the most reliable verification of the pencil-beam algorithm, the treatment planning system's pencil-beam parameters need to be adjusted to better model the electron beam in the penumbra and the dose fall-off region. Because the pencil-beam algorithm does not model all processes involved with electron transport, the pencil-beam algorithm's correction factors for the penumbra regions need to be adjusted in order to perform the calculations correctly [4]. In addition, the other pencil-beam parameters which represent the electron beam's characteristics in the RAHD system need to be verified in a water phantom. Moreover, further investigation is needed in evaluating the performance of the pencil-beam algorithm for three-dimensional dose calculations, which encounter oblique incident surfaces, irregular shapes, and inhomogeneities [1, 6, 7, 10, 14, 16, 18].

6. BIBLIOGRAPHY

- [1] Blomquist, M.; Karlsson, M.; Karlsson M. Test procedures for verification of an electron pencil-beam algorithm implemented for treatment planning. *Radiotherapy and Oncology* 39: 271-286; 1996
- [2] Bova, F. J. A film phantom for routine film dosimetry in the clinical environment. *Medical Dosimetry* 15: 83-85; 1990
- [3] Boyd, R. A.; Hogstrom, K. R.; Starkschall, G. Electron pencil-beam redefinition algorithm dose calculations in the presence of heterogeneities. *Med. Phys.* 28(10): 2096-2104; 2001
- [4] Boyd, R. A.; Hogstrom, K. R.; White, R. A.; Starkschall, G. Modelling pencil-beam divergence with the electron pencil-beam redefinition algorithm. *Phys. Med. Biol.* 46: 2841-2856; 2001
- [5] Cember, H. *Introduction to Health Physics 3rd edition*; McGraw-Hill: 1996
- [6] Craig, T.; Brochu, D.; Dyk, J. V. A quality assurance phantom for three-dimensional radiation treatment planning. *Int. J. Radiation Oncology Biol. Phys.* 44: 955-966; 1999
- [7] Engelsman, M. et al. Impact of simple tissue inhomogeneity correction algorithms on conformal radiotherapy of lung tumours. *Radiotherapy and Oncology* 60: 299-309; 2001
- [8] Evans, M. D. C.; Schreiner, L. J. A simple technique for film dosimetry. *Radiotherapy and Oncology*. 23: 265-267; 1992
- [9] Fujisaki, T. et al. Contribution of Cerenkov radiation in high-energy x-ray and electron beam film dosimetry using water-substitute phantoms. *Phys. Med. Biol.* 48: N105-N109; 2003
- [10] Higgins, P. D.; Gerbi, B. J.; Khan, F. M. Application of measured pencil beam parameters for electron beam model evaluation. *Med. Phys.* 30(4): 514-520; 2003
- [11] Hogstrom, K. R.; Kurup, R. G.; Shiu, A. S.; Starkschall, G. A two-dimensional pencil-beam algorithm for calculation of arc electron dose distribution. *Phys. Med. Biol.* 34: 315-341; 1989
- [12] ICRU. Report 44. Tissue substitutes in radiation dosimetry and measurement. Washington, DC: International Commission on Radiation Units and Measurement; 1989
- [13] Leavitt, D. D. et al. Electron arc therapy: physical measurement and treatment planning techniques. *Int. J. Radiation Oncology Biol. Phys.* 11: 987-999; 1985
- [14] McShan, D. L.; Frass, B. A.; Ten Haken R. K. Dosimetry verification of a 3-D electron pencil beam dose calculation algorithm. *Med. Phys.* 21(1): 13-23; 1994
- [15] Shani, G. *Radiation dosimetry: instrumentation and methods 2nd ed.*; Boca Raton FL: CRC Press: 2001

- [16] Shiu, A. S.; Tung, S.; Hogstrom, K. R. Verification data for electron beam dose algorithms. *Med. Phys.* 19(3): 623-636; 1992
- [17] Thwaites, D. I.; Williams, J. R. *Radiotherapy Physics in practice*; Oxford Medical Publications: 1993
- [18] Wieslander, E.; Knoos, T. A virtual linear accelerator for verification of treatment planning system. *Phys. Med. Biol.* 45: 2887-2896; 2000

7. Appendices

7.1. Appendix A

Optical Density Calibration File with the associated OD versus Scanner A/D
Values plot and the Stouffer Step Wedge Image

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\OD Cal Stepwedge\OD Cal169.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2003/12/16 11:53

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=

PROCESSOR=

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[OD_INFO]

OD_FILENAME=C:\RIT113V4\OD Cal Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD	SCANNER
----	---------

0.070	4095
-------	------

0.160	3157
-------	------

0.270	2350
-------	------

0.380	1754
-------	------

0.500	1241
-------	------

0.630	888
-------	-----

0.740	655
-------	-----

0.840	487
-------	-----

0.970	354
-------	-----

1.080	261
-------	-----

1.200	195
-------	-----

1.320	140
-------	-----

1.440	105
-------	-----

1.550	79
1.690	55
1.790	42
1.920	30
2.010	23
2.160	16
2.290	12
2.410	8
2.530	6




Figure A.1 Plot generated from the OD calibration file by RIT113.v4 software




Figure A.2 Stouffer Step Wedge Film Image has 32 steps (32 shades of gray) – Provided by Radiological Imaging Technology

7.2. Appendix B

Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 150MU _
4cm x 4cm cone _ 6MeV, 9MeV, 12MeV, 16MeV, and 20MeV

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\4x4\6MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/12 9:58

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 591 1.230780e+000

95 709 1.570000e+000

90 815 1.700000e+000

85 934 1.810000e+000

80 1079 1.900000e+000

75 1172 1.980000e+000

70 1315 2.040000e+000

65 1455 2.100000e+000

60 1583 2.160000e+000

55 1682 2.210000e+000

50 1871 2.270000e+000

45 2007 2.330000e+000

40	2211	2.390000e+000
35	2339	2.450000e+000
30	2476	2.520000e+000
25	2659	2.600000e+000
20	2804	2.680000e+000
15	2916	2.760000e+000
10	3028	2.850000e+000
5	3083	2.960000e+000
3.700000e+000	3106	3
6.000000e-001	3151	3.300000e+000
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\4x4\9MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/12 10:10

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	584	1.804020e+000
95	640	2.350000e+000
90	709	2.580000e+000
85	775	2.740000e+000
80	868	2.880000e+000
75	959	3
70	1076	3.130000e+000
65	1180	3.220000e+000
60	1291	3.320000e+000
55	1429	3.400000e+000
50	1618	3.500000e+000
45	1733	3.580000e+000

40	1926	3.670000e+000
35	2075	3.760000e+000
30	2248	3.850000e+000
25	2430	3.940000e+000
20	2563	4.030000e+000
15	2726	4.150000e+000
10	2868	4.290000e+000
5	3017	4.460000e+000
4.200000e+000	3033	4.500000e+000
2.100000e+000	3115	4.700000e+000
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\4x4\12MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/12 10:28

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 498 2.090640e+000

95 570 2.990000e+000

90 632 3.350000e+000

85 703 3.560000e+000

80 759 3.770000e+000

75 835 3.950000e+000

70 962 4.150000e+000

65 1085 4.300000e+000

60 1211 4.460000e+000

55 1317 4.600000e+000

50 1485 4.740000e+000

45 1599 4.850000e+000

40	1764	4.990000e+000
35	1935	5.130000e+000
30	2096	5.250000e+000
25	2255	5.390000e+000
20	2428	5.540000e+000
15	2583	5.680000e+000
10	2663	5.830000e+000
5	2868	6.120000e+000
2	2986	6.500000e+000
1.400000e+000	3039	7
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\4x4\16MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/12 11:27

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=22

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	514	1.921734e+000
95	580	3.480000e+000
90	663	4
85	743	4.300000e+000
80	826	4.600000e+000
75	929	4.850000e+000
70	1051	5.100000e+000
65	1152	5.300000e+000
60	1281	5.530000e+000
55	1438	5.800000e+000
50	1604	6
45	1767	6.200000e+000

40	1923	6.400000e+000
35	2118	6.600000e+000
30	2234	6.770000e+000
25	2383	6.950000e+000
20	2586	7.200000e+000
15	2753	7.400000e+000
10	2872	7.750000e+000
5	2994	8.250000e+000
2.900000e+000		3029 9
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\4x4\20MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/12 11:48

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=21

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	463	1.416014e+000
95	549	3.750000e+000
90	626	4.450000e+000
85	716	4.950000e+000
80	797	5.330000e+000
75	882	5.620000e+000
70	1003	5.950000e+000
65	1137	6.280000e+000
60	1248	6.520000e+000
55	1365	6.800000e+000
50	1526	7.100000e+000
45	1634	7.350000e+000

40	1827	7.650000e+000	
35	1999	7.980000e+000	
30	2149	8.250000e+000	
25	2367	8.600000e+000	
20	2569	8.930000e+000	
15	2724	9.300000e+000	
10	2838	9.760000e+000	
5	2916	1.055000e+001	
0	3.168248e+003		1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

7.3. Appendix C

Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 150MU _
6cm x 6cm cone _ 6MeV, 9MeV, 12MeV, 16MeV, and 20MeV

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\6x6\6MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/12 12: 0

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	614	1.264500e+000
95	709	1.600000e+000
90	797	1.720000e+000
85	922	1.830000e+000
80	1030	1.900000e+000
75	1116	1.980000e+000
70	1268	2.060000e+000
65	1473	2.140000e+000
60	1555	2.200000e+000
55	1766	2.260000e+000
50	1885	2.310000e+000
45	2043	2.360000e+000

40	2149	2.410000e+000
35	2267	2.460000e+000
30	2455	2.530000e+000
25	2599	2.590000e+000
20	2726	2.670000e+000
15	2873	2.750000e+000
10	2989	2.850000e+000
5	3035	2.970000e+000
3.700000e+000	3054	3
7.000000e-001	3119	3.300000e+000
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\6x6\9MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/12 12:36

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=22

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	551	2.022878e+000
95	647	2.550000e+000
90	727	2.730000e+000
85	814	2.870000e+000
80	899	2.990000e+000
75	1015	3.090000e+000
70	1127	3.190000e+000
65	1242	3.270000e+000
60	1379	3.350000e+000
55	1495	3.430000e+000
50	1668	3.520000e+000
45	1820	3.610000e+000

40	2023	3.690000e+000	
35	2179	3.770000e+000	
30	2331	3.840000e+000	
25	2443	3.930000e+000	
20	2675	4.040000e+000	
15	2812	4.160000e+000	
10	2886	4.270000e+000	
5	3021	4.430000e+000	
2	3051	4.700000e+000	
0	3.168248e+003		1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\6x6\12MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/12 12:59

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	503	2.377260e+000
95	667	3.500000e+000
90	754	3.770000e+000
85	847	4
80	976	4.200000e+000
75	1075	4.350000e+000
70	1210	4.480000e+000
65	1311	4.600000e+000
60	1395	4.700000e+000
55	1509	4.800000e+000
50	1617	4.900000e+000
45	1764	5.030000e+000

40	1878	5.130000e+000
35	1999	5.220000e+000
30	2168	5.330000e+000
25	2323	5.480000e+000
20	2484	5.610000e+000
15	2611	5.740000e+000
10	2718	5.900000e+000
5	2855	6.130000e+000
2.100000e+000	3003	6.500000e+000
1.400000e+000	3022	7
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\6x6\16MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/12 13: 6

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=21

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	521	2.461560e+000
95	700	4.250000e+000
90	864	4.800000e+000
85	1013	5.150000e+000
80	1105	5.400000e+000
75	1243	5.600000e+000
70	1349	5.800000e+000
65	1474	5.970000e+000
60	1601	6.150000e+000
55	1731	6.300000e+000
50	1870	6.470000e+000
45	2004	6.620000e+000

40	2141	6.780000e+000	
35	2231	6.900000e+000	
30	2373	7.050000e+000	
25	2501	7.200000e+000	
20	2623	7.400000e+000	
15	2771	7.600000e+000	
10	2841	7.850000e+000	
5	2963	8.270000e+000	
0	3.168248e+003		1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\6x6\20MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/12 13:16

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=21

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	483	1.820880e+000
95	547	4.500000e+000
90	624	5.330000e+000
85	707	5.860000e+000
80	805	6.270000e+000
75	879	6.600000e+000
70	960	6.860000e+000
65	1086	7.140000e+000
60	1217	7.400000e+000
55	1364	7.650000e+000
50	1522	7.900000e+000
45	1678	8.120000e+000

40	1831	8.310000e+000	
35	2012	8.550000e+000	
30	2196	8.800000e+000	
25	2390	9.030000e+000	
20	2535	9.300000e+000	
15	2715	9.600000e+000	
10	2795	9.950000e+000	
5	2880	1.030000e+001	
0	3.168248e+003		1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

7.4. Appendix D

Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 150MU _
10cm x 10cm cone _ 6MeV, 9MeV, 12MeV, 16MeV, and 20MeV

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\10x10\6MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 11:11

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	572	1.331940e+000
95	632	1.600000e+000
90	671	1.720000e+000
85	717	1.810000e+000
80	804	1.900000e+000
75	891	1.970000e+000
70	968	2.030000e+000
65	1069	2.100000e+000
60	1156	2.160000e+000
55	1293	2.230000e+000
50	1445	2.290000e+000
45	1573	2.350000e+000

40	1691	2.400000e+000
35	1819	2.450000e+000
30	2014	2.510000e+000
25	2238	2.600000e+000
20	2388	2.670000e+000
15	2604	2.750000e+000
10	2792	2.850000e+000
5	2949	2.950000e+000
3.400000e+000	2994	3
6.000000e-001	3129	3.300000e+000
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD_SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\10x10\9MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 11:23

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=24

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	537	2.090640e+000
95	606	2.510000e+000
90	693	2.710000e+000
85	774	2.850000e+000
80	850	2.970000e+000
75	937	3.070000e+000
70	1042	3.150000e+000
65	1171	3.250000e+000
60	1322	3.350000e+000
55	1438	3.420000e+000
50	1582	3.500000e+000
45	1711	3.580000e+000

40	1884	3.660000e+000
35	2029	3.730000e+000
30	2253	3.820000e+000
25	2423	3.920000e+000
20	2557	4.020000e+000
15	2710	4.130000e+000
10	2850	4.240000e+000
5	2994	4.430000e+000
4	3055	4.500000e+000
1.900000e+000	3101	4.700000e+000
1	3114	5
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\10x10\12MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 11:31

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	525	2.748180e+000
95	605	3.550000e+000
90	664	3.830000e+000
85	725	4.020000e+000
80	813	4.200000e+000
75	907	4.330000e+000
70	1002	4.470000e+000
65	1115	4.590000e+000
60	1250	4.700000e+000
55	1354	4.800000e+000
50	1491	4.900000e+000
45	1619	5

40	1790	5.130000e+000
35	1933	5.230000e+000
30	2132	5.350000e+000
25	2247	5.450000e+000
20	2476	5.580000e+000
15	2591	5.700000e+000
10	2787	5.870000e+000
5	2950	6.140000e+000
2.200000e+000	3041	6.500000e+000
1.500000e+000	3045	7
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\10x10\16MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 11:41

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=22

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 500 3.102240e+000

95 600 4.620000e+000

90 676 5.050000e+000

85 746 5.310000e+000

80 858 5.560000e+000

75 973 5.780000e+000

70 1041 5.920000e+000

65 1165 6.100000e+000

60 1292 6.250000e+000

55 1444 6.400000e+000

50 1578 6.550000e+000

45 1714 6.670000e+000

40	1847	6.800000e+000	
35	2004	6.950000e+000	
30	2188	7.100000e+000	
25	2331	7.250000e+000	
20	2501	7.400000e+000	
15	2655	7.600000e+000	
10	2810	7.800000e+000	
5	2938	8.250000e+000	
3	2988	9	
0	3.168248e+003		1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\10x10\20MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 11:51

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=21

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 498 2.073780e+000

95 576 5.050000e+000

90 653 5.850000e+000

85 720 6.320000e+000

80 811 6.700000e+000

75 914 7

70 1006 7.250000e+000

65 1132 7.530000e+000

60 1252 7.760000e+000

55 1411 7.980000e+000

50 1567 8.170000e+000

45 1686 8.350000e+000

40	1830	8.530000e+000	
35	2007	8.730000e+000	
30	2172	8.920000e+000	
25	2340	9.170000e+000	
20	2536	9.430000e+000	
15	2651	9.650000e+000	
10	2798	9.950000e+000	
5	2946	1.092000e+001	
0	3.168248e+003		1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

7.5. Appendix E

Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 150MU _
15cm x 15cm cone _ 6MeV, 9MeV, 12MeV, 16MeV, and 20MeV

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 03-Mar-04\15x15\6MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 10: 8

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	558	1.298220e+000
95	586	1.530000e+000
90	663	1.680000e+000
85	710	1.760000e+000
80	804	1.850000e+000
75	882	1.930000e+000
70	980	2
65	1096	2.070000e+000
60	1186	2.130000e+000
55	1298	2.180000e+000
50	1462	2.240000e+000
45	1586	2.300000e+000

40	1760	2.360000e+000
35	1926	2.420000e+000
30	2063	2.480000e+000
25	2231	2.550000e+000
20	2447	2.620000e+000
15	2635	2.700000e+000
10	2779	2.790000e+000
5	2936	2.900000e+000
3	3035	3
6.000000e-001	3109	3.300000e+000
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 03-Mar-04\15x15\9MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 10: 0

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 534 2.090640e+000

95 594 2.490000e+000

90 679 2.680000e+000

85 758 2.820000e+000

80 839 2.930000e+000

75 915 3.030000e+000

70 1028 3.130000e+000

65 1151 3.230000e+000

60 1251 3.300000e+000

55 1409 3.400000e+000

50 1541 3.480000e+000

45 1671 3.560000e+000

40	1867	3.640000e+000
35	2031	3.720000e+000
30	2199	3.810000e+000
25	2375	3.900000e+000
20	2538	3.990000e+000
15	2678	4.100000e+000
10	2852	4.220000e+000
5	2977	4.410000e+000
3.600000e+000	3073	4.500000e+000
8.000000e-001	3118	5
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 03-Mar-04\15x15\12MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 10:27

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=22

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 515 2.815620e+000

95 611 3.540000e+000

90 697 3.800000e+000

85 772 4

80 898 4.180000e+000

75 1007 4.330000e+000

70 1138 4.460000e+000

65 1241 4.580000e+000

60 1358 4.690000e+000

55 1529 4.800000e+000

50 1647 4.890000e+000

45 1828 5

40	1943	5.110000e+000	
35	2117	5.210000e+000	
30	2305	5.330000e+000	
25	2454	5.440000e+000	
20	2570	5.550000e+000	
15	2706	5.690000e+000	
10	2836	5.840000e+000	
5	2989	6.100000e+000	
2	3049	6.500000e+000	
0	3.168248e+003		1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\15x15\16MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 10:40

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 513 3.169680e+000

95 619 4.580000e+000

90 693 5

85 760 5.310000e+000

80 893 5.560000e+000

75 1013 5.760000e+000

70 1119 5.930000e+000

65 1253 6.100000e+000

60 1388 6.250000e+000

55 1477 6.370000e+000

50 1635 6.500000e+000

45 1801 6.660000e+000

40	1967	6.800000e+000
35	2141	6.950000e+000
30	2306	7.100000e+000
25	2428	7.230000e+000
20	2609	7.400000e+000
15	2724	7.600000e+000
10	2851	7.800000e+000
5	2991	8.250000e+000
3.100000e+000	3034	9
2.700000e+000	3043	1.050000e+001
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\15x15\20MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 10:59

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=21

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 519 2.090640e+000

95 603 5.250000e+000

90 648 5.980000e+000

85 715 6.450000e+000

80 795 6.800000e+000

75 858 7.070000e+000

70 949 7.300000e+000

65 1071 7.530000e+000

60 1208 7.800000e+000

55 1301 7.980000e+000

50 1418 8.200000e+000

45 1580 8.400000e+000

40	1723	8.550000e+000	
35	1868	8.750000e+000	
30	2068	8.950000e+000	
25	2273	9.200000e+000	
20	2427	9.400000e+000	
15	2595	9.700000e+000	
10	2735	10	
5	2912	11	
0	3.168248e+003		1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD_SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

7.6. Appendix F

Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 150MU _
20cm x 20cm cone _ 6MeV, 9MeV, 12MeV, 16MeV, and 20MeV

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\20x20\6MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 15:31

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 530 1.197060e+000

95 577 1.550000e+000

90 623 1.680000e+000

85 714 1.780000e+000

80 770 1.860000e+000

75 857 1.940000e+000

70 942 2

65 1052 2.070000e+000

60 1135 2.130000e+000

55 1298 2.190000e+000

50 1411 2.250000e+000

45 1559 2.310000e+000

40	1742	2.370000e+000
35	1871	2.420000e+000
30	2015	2.480000e+000
25	2206	2.550000e+000
20	2415	2.630000e+000
15	2607	2.720000e+000
10	2777	2.800000e+000
5	2929	2.900000e+000
3	3022	3
6.000000e-001	3119	3.300000e+000
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal
 OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\20x20\9MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 15:52

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=24

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100	525	1.955760e+000
95	621	2.460000e+000
90	721	2.700000e+000
85	786	2.830000e+000
80	884	2.950000e+000
75	985	3.050000e+000
70	1099	3.150000e+000
65	1226	3.240000e+000
60	1358	3.330000e+000
55	1519	3.400000e+000
50	1676	3.490000e+000
45	1802	3.580000e+000

40	1999	3.660000e+000
35	2120	3.730000e+000
30	2278	3.820000e+000
25	2429	3.900000e+000
20	2585	3.990000e+000
15	2707	4.090000e+000
10	2844	4.210000e+000
5	2996	4.400000e+000
3.800000e+000	3038	4.500000e+000
1.700000e+000	3050	4.700000e+000
8.000000e-001	3081	5
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\20x20\12MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 16:12

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 536 2.680740e+000

95 634 3.500000e+000

90 714 3.820000e+000

85 796 4

80 889 4.180000e+000

75 990 4.320000e+000

70 1095 4.460000e+000

65 1216 4.580000e+000

60 1360 4.690000e+000

55 1484 4.800000e+000

50 1673 4.900000e+000

45 1797 5.010000e+000

40	1941	5.110000e+000	
35	2079	5.200000e+000	
30	2271	5.330000e+000	
25	2429	5.440000e+000	
20	2575	5.570000e+000	
15	2702	5.700000e+000	
10	2848	5.860000e+000	
5	2963	6.130000e+000	
2.200000e+000	3031	6.500000e+000	
1.600000e+000	3062	7	
0	3.168248e+003	1000	

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\20x20\16MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 16:23

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 515 3.220260e+000

95 598 4.630000e+000

90 646 5.060000e+000

85 732 5.350000e+000

80 806 5.550000e+000

75 873 5.750000e+000

70 970 5.930000e+000

65 1100 6.100000e+000

60 1212 6.250000e+000

55 1330 6.400000e+000

50 1403 6.530000e+000

45 1580 6.660000e+000

40	1694	6.800000e+000
35	1893	6.970000e+000
30	2044	7.100000e+000
25	2214	7.250000e+000
20	2387	7.400000e+000
15	2495	7.550000e+000
10	2700	7.800000e+000
5	2894	8.350000e+000
3.200000e+000	2997	9
2.900000e+000	3007	1.050000e+001
0	3.168248e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal
 OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 150MU 03-Mar-04\20x20\20MeV.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 3/11 16:38

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=21

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 486 2.613300e+000

95 609 5.240000e+000

90 682 5.980000e+000

85 753 6.450000e+000

80 848 6.800000e+000

75 926 7.120000e+000

70 1039 7.360000e+000

65 1135 7.590000e+000

60 1263 7.820000e+000

55 1380 8

50 1542 8.200000e+000

45 1694 8.400000e+000

40	1819	8.550000e+000	
35	2023	8.780000e+000	
30	2205	9	
25	2347	9.200000e+000	
20	2516	9.430000e+000	
15	2663	9.700000e+000	
10	2788	10	
5	2910	11	
0	3.168248e+003		1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

7.7. Appendix G

Parallel dose calibration file using Kodak Ready-Pack EDR-2 film _ 100MU_
15cm x 15cm cone _ 12MeV

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\EDR films Jan-18-04\Calibration\12MeV15x15.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 2/18 18: 7

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=23

INTERPOLATION=cubic

FILM=EDR2

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 1130 2.730885e+000

95 1236 3.510000e+000

90 1355 3.800000e+000

85 1427 4

80 1531 4.180000e+000

75 1628 4.330000e+000

70 1764 4.490000e+000

65 1884 4.600000e+000

60 1956 4.700000e+000

55 2086 4.800000e+000

50 2196 4.900000e+000

45 2303 5

40	2435	5.150000e+000	
35	2517	5.250000e+000	
30	2727	5.450000e+000	
25	2757	5.480000e+000	
20	2836	5.550000e+000	
15	2915	5.680000e+000	
10	2989	5.800000e+000	
		6.400000e+000	3070 6
2	3123	6.500000e+000	
		1.400000e+000	3158 7
0		3.164963e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD_SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

Appendix H

Parallel dose calibration file using Kodak Ready-Pack XV-2 film _ 100MU_
15cm x 15cm cone _ 12MeV

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Cal Parallel 16-Jan-04\12MeV\XV 99cGy.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 1/16 13: 9

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION= 85

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=16

INTERPOLATION=cubic

FILM=XV

[CALIBRATION_DATA]

DOSE	SCANNER	LOCATION
9.930000e+001	7	2.930620e+000
99	8	3
9.702000e+001	10	3.300000e+000
9.334000e+001	11	3.600000e+000
8.957000e+001	13	3.800000e+000
8.699000e+001	14	3.900000e+000
8.421000e+001	15	4
8.162000e+001	17	4.100000e+000
7.845000e+001	19	4.200000e+000
6.752000e+001	34	4.500000e+000
4.439000e+001	125	5
2.204000e+001	626	5.500000e+000

6.360000e+000	1845	6
1.990000e+000	2501	6.500000e+000
1.390000e+000	2646	7
0	3.293994e+003	1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal85.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3148
2.700000e-001	2344
3.800000e-001	1750
5.000000e-001	1237
6.300000e-001	882
7.400000e-001	648
8.400000e-001	484
9.700000e-001	351
1.080000e+000	258
1.200000e+000	194
1.320000e+000	140
1.440000e+000	104
1.550000e+000	80
1.690000e+000	56
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

7.9. Appendix I

Parallel dose calibration file using Kodak Ready-Pack XV-2 film _ 50MU_ 10cm
x 10cm cone _ 12MeV

[RIT113_V4_CAL]

FILE_NAME=C:\RIT113V4\Tom's Films\Calibration\12MeV60MU10x10.cal

WARNING: Do not modify section headings or keywords in this file. Unexpected results may occur.

WARNING: Modifying calibration data in this file will change your dosimetric results

DATE_CREATED = 2004/ 2/12 10:31

[FACILITY_INFO]

FACILITY_NAME=Samaritan Regional Cancer Center

LINAC=LINAC1

PROCESSOR=Processor1

COMMENTS=

[SCANNER_INFO]

SCANNER_SERIAL= 0

INTEGRATION_TIME=12.0

DIGITIZER=VXR12

FIRMWARE=5.05

TOOLBOX=3.1.9.3

SCANNING_RESOLUTION=169

SCAN_BITS=12

LUT=0

[CALIBRATION_INFO]

NUMBER_POINTS=14

INTERPOLATION=cubic

FILM=XV

[CALIBRATION_DATA]

DOSE SCANNER LOCATION

100 59 2.731320e+000

95 65 3.550000e+000

90 76 3.850000e+000

85 84 4

80 100 4.200000e+000

70 140 4.480000e+000

60 199 4.700000e+000

50 288 4.900000e+000

40 427 5.100000e+000

30 721 5.350000e+000

20 1108 5.550000e+000

10 1996 5.900000e+000

5 2626 6.200000e+000
 0 3.354395e+003 1000

[OD_INFO]

OD_FILENAME=C:\RIT113V4\Cal OD Stepwedge\OD Cal169.cal

OD_NUMBER_POINTS=22

[OD_DATA]

OD SCANNER

7.000000e-002	4095
1.600000e-001	3157
2.700000e-001	2350
3.800000e-001	1754
5.000000e-001	1241
6.300000e-001	888
7.400000e-001	655
8.400000e-001	487
9.700000e-001	354
1.080000e+000	261
1.200000e+000	195
1.320000e+000	140
1.440000e+000	105
1.550000e+000	79
1.690000e+000	55
1.790000e+000	42
1.920000e+000	30
2.010000e+000	23
2.160000e+000	16
2.290000e+000	12
2.410000e+000	8
2.530000e+000	6

Appendix J

Film images and iso-dose contours on Kodak Ready-Pack EDR-2 film using
150M_ 15cm x 15cm cone_ 6MeV, 9MeV, 12MeV, 16MeV, and 20MeV

Figure J.1 6MeV 15x15 cone 150MU EDR film





Figure J.2 9MeV 15x15 cone 150MU EDR film





Figure J.3 12MeV 15x15 cone 150MU EDR film





Figure J.4 16MeV 15x15 cone 150MU EDR film





Figure J.6 20MeV 15x15 cone 150MU EDR film

